

**ANALOGUE OUTPUT SIGNALS**

OUT.CLAMP	OUTPUT	
12 13	1	ClO <sub>2</sub> 1
14 15	2	ClO <sub>2</sub> 2 / pH
16 17	3	AR.AVERAGE/CONTROLLER 2
18 19	4	CONTROLLER 1


**ANALYZER-PROBE CONNECTIONS, CHLORINE DIOXIDE MEASUREMENT**

INPUT CLAMP	MPC66	PROBE	
20	INPUT 1	20	ClO <sub>2</sub> 1
21	INPUT 1	19	ClO <sub>2</sub> 1
22	INPUT 2	20	ClO <sub>2</sub> 2
23	INPUT 2	19	ClO <sub>2</sub> 2


**ANALYZER-PROBE CONNECTIONS, CHLORINE DIOXIDE AND pH MEASUREMENT**

INPUT CLAMP	MPC66	PROBE	
20	INPUT 1	20	ClO <sub>2</sub> 1
21	INPUT 1	19	ClO <sub>2</sub> 1
22	INPUT 2	20	pH
23	INPUT 2	19	pH
6	CLEANING	17a	CLEANING
7	CLEANING	18a	CLEANING

**CHLORINE DIOXIDE MEASUREMENT**



**PH MEASUREMENT (TEMPERATURE USED FOR COMPENSATION ONLY)**



z	GREEN
č	BLACK
b	WHITE
r	RED
me	MAIN ELECTRODE
re	REFERENCE ELECTRODE